

Thank you for your inquiry regarding Beekley Medical® SPOT® skin markers.

All Beekley Medical® SPOT® products are durable hypoallergenic, latex-free devices and do not contain either natural rubber latex or dry natural rubber as a component in the product or in packaging we supply it in.

The adhesive on Beekley Medical® SPOT® is a hypo-allergenic, medical-grade, skin contact pressure sensitive acrylic.

It was tested to be non-irritating in both animals and humans. The adhesive is similar to the type you might find on an over-the-counter band-aid type of product. Testing was completed by the manufacturer in accordance with the ISO 10993 Part - 1 “Biological Evaluation of Medical Devices,” as put forth by the FDA.

The adhesive testing that was conducted on humans satisfied FDA requirements to earn the hypoallergenic rating. The exact ingredients of the adhesive are proprietary and not published.

Should you have any questions or need further assistance, please do not hesitate to contact us at 1.800.BEEKLEY (233.5539).